

(IgG)

Sample No.	Elegance	Sero-CP	Micro-IF		Western Blot	
			C.bn	C.t	C.bn	C.t
1	0.89	1.75	16	128<	-	+
2	0.65	1.34	<8	128<	-	+
3	0.84	1.29	32	128<	-	+
4	0.69	0.91	8	128<	-	+
5	0.46	1.33	32	128<	-	+
6	0.76	1.06	16	128<	-	+
7	0.38	1.22	<8	128<	-	+
8	0.62	0.99	<8	128<	-	+
9	0.75	1.36	<8	32	-	-
10	0.37	1.23	<8	32	-	-
11	1.08	1.87	<8	128<	-	-
12	0.84	0.95	<8	64	-	-
13	0.26	0.55	<8	<8	-	-
14	0.64	0.86	8	8	-	-
15	0.24	0.57	<8	<8	-	-
16	0.38	0.57	<8	128<	-	+
17	0.76	1.03	<8	128<	-	-
18	0.37	0.57	<8	8	-	-
19	0.37	0.71	8	8	-	-
20	3.45	3.25	128<	128<	+	+
21	1.63	1.53	128<	128<	+	+
22	2.52	1.79	128<	128<	+	+
23	2.44	1.94	128<	128<	+	+
24	1.50	1.51	128<	128<	+	+
25	2.54	2.32	128<	64	+	+
26	3.13	2.71	128<	32	+	+
27	3.27	3.00	128<	64	+	+
28	1.99	2.18	64	32	+	+
29	3.25	2.50	128<	128<	+	+
30	2.52	2.90	128<	32	+	+
31	0.06	0.28	<8	<8	-	-
32	1.36	1.22	64	<8	+	-
33	0.40	0.82	<8	<8	-	-
34	0.06	0.38	<8	<8	-	-
35	0.35	0.54	<8	<8	-	-
36	1.66	1.93	64	<8	+	-
37	1.86	1.88	64	<8	+	-
38	1.44	1.40	64	<8	+	-
39	3.05	2.79	128<	<8	+	-
40	2.58	2.48	128<	<8	+	-
41	3.02	2.64	128<	<8	+	-
42	2.13	2.00	128<	<8	+	-
43	1.32	1.54	128<	<8	+	-
44	2.68	2.27	128<	<8	+	-
45	2.21	2.29	128<	8	+	-
46	1.29	1.47	64	<8	+	-
47	5.00	3.57	128<	<8	+	-

(IgA)

Sample No.	Elegance	Sero-CP	Micro-IF		Western Blot	
			C.bn	C.t	C.bn	C.t
1	0.60	1.06	64	128<	-	+
2	0.29	0.54	<8	32	-	+
3	0.26	0.71	<8	128<	-	+
4	0.44	0.60	<8	<8	-	+
5	0.21	0.49	<8	<8	-	+
6	0.24	0.62	<8	64	-	+
7	0.26	0.65	<8	32	-	+
8	1.02	1.04	8	32	-	+
9	0.67	4.55	<8	32	-	-
10	0.65	0.43	<8	<8	-	-
11	2.60	2.37	<8	32	+	+
12	1.93	1.31	<8	<8	+	-
13	0.62	0.75	<8	<8	+	-
14	1.99	5.37	<8	8	+	-
15	0.19	0.44	<8	<8	-	-
16	0.41	0.69	<8	64	-	+
17	0.66	0.64	<8	<8	-	-
18	0.81	0.96	<8	<8	+	-
19	1.25	0.93	<8	<8	-	-
20	0.65	1.87	<8	128<	-	+
21	1.37	0.73	<8	128<	+	+
22	0.85	0.94	<8	128<	+	+
23	0.81	1.09	<8	128<	+	+
24	2.23	1.38	<8	64	+	+
25	2.00	2.03	128<	32	+	+
26	1.66	1.76	<8	32	+	+
27	2.47	2.85	<8	32	+	+
28	0.97	0.87	8	32	+	+
29	2.24	2.06	8	32	+	+
30	1.16	2.20	32	16	+	-
31	0.14	0.28	<8	<8	-	-
32	0.86	0.54	<8	<8	-	-
33	0.17	0.39	<8	<8	-	-
34	0.11	0.34	<8	<8	-	-
35	0.29	0.48	<8	<8	-	-
36	1.75	1.16	<8	<8	+	-
37	1.42	0.84	n.d.	n.d.	n.d.	n.d.
38	1.16	1.01	32	<8	+	-
39	1.48	1.18	128<	8	+	-
40	1.64	1.66	128<	<8	+	-
41	1.57	3.21	128<	<8	+	-
42	2.45	1.67	128<	<8	+	-
43	3.23	1.99	<8	<8	+	-
44	1.76	1.14	<8	<8	+	-
45	2.89	2.47	16	<8	+	-
46	1.84	0.83	<8	<8	+	-
47	0.97	2.09	<8	<8	+	-

Bioclone's Quality Management System certified to ISO 9001, ISO 13485, GMP and CE Mark

Bioclone Australia Pty Limited (a subsidiary of Hitachi Chemical Co., Ltd) ABN 14 002 036 071

71 - 73 Railway Parade, Marrickville, Sydney NSW 2204 AUSTRALIA

Tel: +61 (0) 2 9517 1966 Freecall 1800 251 138 Fax: +61 (0) 2 9517 2990

E-mail: sales@bioclone.com.au Web: www.bioclone.com.au

Form # CPCOMP Pt2 Ed 1